

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

JAN 14 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130571</u>	
Name <u>Patrick E. Martin</u>		ID# <u>4072</u>	Date <u>1-11-10</u> ^{12N80}
A	Agency <u>APD</u>	Phone # <u>764-9592</u>	
Instrument Location <u>Seely Sub</u>			
Alco S/N <u>X172931</u>		Target Value <u>.082</u>	High Pressure <u>600</u>
B	Alco Test Value Average <u>.082</u> <u>.083</u> 1 st Alco 2 nd Alco		
Signature <u>P E Martin</u>		<u>083</u> <u>1/22/10</u>	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of January, 2010

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

JANUARY 12, 2010

OPERATOR'S NAME:

MARTIN, PATRICK E

OPERATOR'S NUMBER: 4072

SUBJECT'S LAST NAME:

MARTIN

SUBJECT'S FIRST NAME/MI :

PATRICK

O.L. #: 1234567

DEPT/AGENCY: ANC1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .082

ALCO S/N: X172931

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.36 in		
ALCO TARGET	.080	10:53
BLANK TEST	.000	10:54
INTERNAL STANDARD	VERIFIED	10:54
ALCO TV 29.36 in	.082	10:54
BLANK TEST	.000	10:55
SUBJECT SAMPLE	.000	10:55
BLANK TEST	.000	10:56
ALCO TV 29.36 in	.083	10:56
BLANK TEST	.000	10:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

JANUARY 12, 2010

TIME 10:51

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	51c
BREATH TUBE:	41c
BAROMETER:	29.36 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~`